

Background

- Metagenomics provides an opportunity to understand the microbial population present in a given environment.
- The development of high-throughput sequencing has made the study of microbiomes increasingly possible.
- However, with recent increased activity in metagenomics research, there is need for reference materials that enable data accuracy and quality to be assessed.
- Control materials could enable performance evaluation of sample processing, library preparation, sequencing methods, and data analysis, thus aiding in the comparison of different studies.

Materials and Methods

We developed a metagenomic control material (MCM) mock community comprising genomic DNA prepared from 10 pathogenic bacterial species mixed at varying concentrations (Table 1, Figure 1). The batch-to-batch production of this material was characterized using digital PCR (Figure 2). This reference material will be available from ATCC under the catalog number ATCC[®] MSA-4000[™].

Using this material, we investigated various methods of microbial genome quantification, including absorbance, fluorescence, high-throughput qPCR (ht-qPCR), digital PCR (dPCR), different 16S rRNA amplicon sequencing strategies, and whole genome sequencing (WGS) (Figure 3).

Table 1. Individual bacterial strains within the MCM genomic DNA mock community

ATCC [®] No.	Species	Abundance	Gram Status	% GC	Genome Size (Mb)	GenBank ID	16S copies
17978 [™]	<i>Acinetobacter baumannii</i>	0.10%	-	39	4	CP000521	5
700802 [™]	<i>Enterococcus faecalis</i>	0.70%	+	37.3	3.34	AE016830	4
700928 [™]	<i>Escherichia coli</i>	1.40%	-	50.6	5.23	AE014075	7
700721 [™]	<i>Klebsiella pneumoniae</i>	14.40%	-	57.1	5.32	CP000647	8
700532 [™]	<i>Neisseria meningitidis</i>	28.90%	-	51.7	2.19	AM421808	4
47085 [™]	<i>Pseudomonas aeruginosa</i>	0.30%	-	66.6	6.26	AE004091	4
BAA-1556 [™]	<i>Staphylococcus aureus</i> (MRSA)	0.70%	+	32.8	2.87	CP000255	5
BAA-1718 [™]	<i>Staphylococcus aureus</i> (MSSA)	14.40%	+	32.8	2.87	AASB02000000	5
BAA-611 [™]	<i>Streptococcus agalactiae</i>	2.90%	+	35.6	2.16	AE009948	7
700669 [™]	<i>Streptococcus pneumoniae</i>	28.90%	+	39.6	2.22	FM211187	4
700294 [™]	<i>Streptococcus pyogenes</i>	7.20%	+	38.5	1.85	AE004092	6




Figure 1. Mass-based composition of MCM




Figure 2. Batch-to-batch production. Evaluation of the composition of two different MCM mock community batches using dPCR (error bars: 95% confidence interval). dPCR enabled the calibration-free value assignment and comparison of different batches.

Results

- Control materials enabled different measurement approaches to be assessed.
- All methods gave comparable results; however, the MCM mock community was able to demonstrate where and by what magnitude the methods differed.
- The ht-qPCR results agreed with the Qubit[®] (Thermo Fisher Scientific) results; the standard curves were generated using calibrators initially quantified by fluorescence.
- The most significant difference observed between the various methods was the relative abundance of *Pseudomonas aeruginosa*.




Figure 3. Assessment of the MCM mock community using different methods

Conclusions

- We demonstrated reproducible production of the MCM mock community.
- There was good agreement between the methods investigated with < 2.5 fold difference in the MCM mock community composition.
- These findings demonstrated that the MCM mock community can assist in evaluating the technical performance of different molecular quantification methods frequently used in microbiome analyses.
- Further application would allow laboratories to monitor results and compare technical performance with other laboratories, enabling them to identify sources of error that will improve accuracy and comparability when performing microbiome analysis as well as pathogen detection using molecular methods.

References

1. O'Sullivan DM, *et al.* Assessing the Accuracy of Quantitative Molecular Microbial Profiling. *Int J Mol Sci* 15(11): 21476-21491, 2014.
2. Huggett JF, *et al.* Considerations for the Development and Application of Control Materials to Improve Metagenomic Microbial Community Profiling. *Accred Qual Assur* 18(2): 77-83, 2013.

Acknowledgements

We would like to acknowledge Dr. Thomas Laver and Dr. David Studholme from the University of Exeter, and Professor Philip Butcher and Dr. Ken Laing from St. Georges, University of London.

Disclaimers

© 2017 American Type Culture Collection. The trademark and trade name, and any other trademarks listed in this publication are trademarks owned by the American Type Culture Collection unless indicated otherwise. © LGC Limited, 2017. All rights reserved. Qubit[®] is a registered trademark of Thermo Fisher Scientific Inc.

The work described here was partially funded by the UK Government Department for Business, Energy & Industrial Strategy (BEIS) and the European Metrology Research Programme (EMRP) joint research project [HLT08] "Infect-Met" (infectmet.lgcgroup.com), which is jointly funded by the EMRP participating countries within EURAMET and the European Union.